



Confluent Medical Technologies
47533 Westinghouse Drive, Fremont, CA 994539
Phone: (510) 683-2000 Fax: (510) 683-2001
Web site: www.confluentmedical.com

March 27, 2019

To Whom It May Concern:

Subject: RoHS 3 DIRECTIVE 2015/863 SUPPLIER DECLARATION OF COMPLIANCE

All Confluent Medical Technologies products consisting of only Nitinol conform to the Directive (EU) 2015/863 amendment of the RoHS DIRECTIVE 2011/65/EU: and do not contain any of the below mentioned 10 Hazardous substances above the specified limits.

- Lead (Pb): < 1000 ppm
- Mercury (Hg): < 100 ppm
- Cadmium (Cd): < 100 ppm
- Hexavalent Chromium: (Cr VI) < 1000 ppm
- Polybrominated Biphenyls (PBB): < 1000 ppm
- Polybrominated Diphenyl Ethers (PBDE): < 1000 ppm
- Bis(2-Ethylhexyl) phthalate (DEHP): < 1000 ppm
- Benzyl butyl phthalate (BBP): < 1000 ppm
- Dibutyl phthalate (DBP): < 1000 ppm
- Diisobutyl phthalate (DIBP): < 1000 ppm

If I can be of further assistance or require additional information, please do not hesitate to call me at (401) 562-2815.

Sincerely;


Jean E. Mattar
Sr. Director QA RA & Compliance
Confluent Medical Technologies

Email: jean.mattar@confluentmedical.com